

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

JUN 23 2009

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130117 ✓
A	Name	Brian Mitchell	ID# 4119 Date 6-21-09
	Agency	Kodiak Police Department	Phone # 486-8000
	Instrument Location	Booking room.	✓
B	Alco S/N	X172991 ✓	Target Value .078 High Pressure 600
	Alco Test Value Average	.075 ✓	.075 ✓
		1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
	Signature	[Signature]	
		(OVER)	

BMB 6/25/09

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz

Scientific Director


State Breath Alcohol Program

Subscribed and sworn before me this 20<sup>th</sup> day of June, 2009.

Carolyn M. Noland

Notary Public, State of Alaska

Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130117

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130117

JUNE 21, 2009

OPERATOR'S NAME:

SUPERVISOR

OPERATOR'S NUMBER: 4119

SUBJECT'S LAST NAME:

VERIFICATION

SUBJECT'S FIRST NAME/MI :

VERIFICATION

O.L. #: XXX

DEPT/AGENCY: KOD1

CASE/REPORT: XXX

TEST TYPE: U

ALCO TARGET VALUE: .078

ALCO S/N: X172991

### — BREATH ANALYSIS —

.078 ADJUSTED FOR 29.77 in  
ALCO TARGET .077 21:59  
BLANK TEST .000 22:00  
INTERNAL STANDARD VERIFIED 22:00  
ALCO TU 29.77 in .075 22:01  
BLANK TEST .000 22:01  
SUBJECT SAMPLE .000 22:02  
BLANK TEST .000 22:03  
ALCO TU 29.77 in .075 22:03  
BLANK TEST .000 22:04

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130117

JUNE 21, 2009  
TIME 22:07

### — DIAGNOSTIC CHECK —

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 50c  
BREATH TUBE: 43c  
BAROMETER: 29.77 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHI  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefghijklmnopqrstuvwxyz{|}~